Name of Cleaning, Disinfection or Sterilization	3M Rapid Multi Enzyme Cleaner	CIDEZYME/ ENZOL	EMpower	Klenzyme	Metrizyme	Prolystica 2x conc. Enzymatic Presoak & cleaner	Transeptic	CaviWipes	CIDEX	Cidex ADS	CIDEX OPA	Cidex Plus	DETERGICIDE	Dynacide PA	EndoAct	EndoDis	HYAMINE	METRICIDE	METRICIDE 28	Metricide Plus 30	Metricide OPA Plus	NU-Cidex	Omnicide 14 N.S.	Omnicide 28	Oxivir Wipes	PI-SPRAY	PI-SPRAY II	PROTEX Disinfection Spary & Wipes	PROTEX ULTRA	Revital-Ox Resert XL HDL	Sani-Cloth Active Wipes	Sanicloth HB	Sani-Cloth Plus	SONO ULTRASOUND WIPES	Super Sani-Cloth Germicidal Wipe	TD100 & TD5	Trophon EPR	T-SPRAY	T-SPRAY II	Virkon	Virkon S	Wex-Cide 128	PERASAFE	Sporicidin	WAVICIDE-01	Autoclave (Steam)	EO sterilization	STERIS SYSTEM1	STERIS V-Pro	STERIS V-Pro Applicable SN	Sterrad 50/ 100S/ 200	Sterrad NX/ 100NX
C22I	C	C					N		C		C	C	C	C	C*7	C*7	C							C	C	C	C	C	C	C	C	C	C		C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
C22K		C							C		C	C	C		C	C	C											C	C		C	C	C		C		C								N	C*5		C*4		C	C
C22P		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C			C				C	N	C*5		C*4		C	C
C22T	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
C25		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C							C	N	C*5		C*4		C	C
C251		C							C		C	C	C		C	C	C	C	C						C				C	N	C				C						C			C	N	C*5	N	C*4	C from first SN	C	C
C252	C	C					N		C		C	C	C	C	C	C	C	C	C						C	C	C	C	C	N	C	C	C	C	C		C	C	C			C	C	C	N	C*5	N	C*4	C from first SN	C	C
C253	C	C					N		C		C	C	C	C	C	C	C								C	C	C	C	C	N	C	C	C	C	C		N	C	C			C	C	C	N	C*5	N	C*4	M00201～	C	C
C25P		C							C		C	C	C		C	C	C	C	C						C				C	N	C				C		C				C			C	N	C*5	N	C*4	C from first SN	C	C
C35	C	C					N		C		C	C	C	C	C	C	C	C	C						C	C	C	C	C	N	C	C	C	C	C		C	C	C			C	C	C	N	C*5	N	C*4	C from first SN	C	C
C41		C							C		C	C	C		C	C	C											C	C		C	C	C		C		C								N	C*5		C*4		C	C
C41B		C							C		C	C	C		C*7	C*7	C	C	C						C				C		C				C		C*8			C		C			N	C*5	C*7	C*4	C from first SN	C	C
C41L47RP		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		N								N	C*5		C*4		C	C
C41RP	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
C41V		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C			C				C	N	C*5		C*4		C	C
C41V1		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C			C				C	N	C*5		C*4		C	C
C42		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C			C				C	N	C*5		C*4		C	C
C42K		C							C		C	C	C		C	C	C											C	C		C	C	C		C		C								N	C*5		C*4		C	C
C42T	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C		C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
CC41R		C							C		C	C	C		C*7	C*7	C	C	C						C				C	N	C				C		C				C			C	N	C*5	C*7	C*4	C from first SN	C	C
CC41R1	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
CL4416R	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
L34		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C			C				C	N	C*5		C*4		C	C
L43K		C					N				C	C			C*7	C*7	C											C	C					C	C						C			C	N	C*5	C*7	C*4	C from first SN	C	C
L44		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C			C				C	N	C*5		C*4		C	C
L441		C					N		C		C	C	C		C	C	C									C	C	C	C	N					C		C	C	C			C	C	C	N	N	N	C*4	C from first SN	C	C
L442	C	C					N		C		C	C	C	C	C	C	C								C	C	C	C	C	N	C	C	C	C	C		N	C	C			C	C	C	N	C*5	N	C*4	M00201～	C	C
L44K	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
L44LA		C							C		C																										N										N	C*5		C*4	C from first SN	C	C
L44LA1	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		N	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
L46K	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
L46K1	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
L51K		C					N				C	C			C*7	C*7	C											C	C					C	C						C			C	N	C*5	C*7	C*4	C from first SN	C	C
L53K	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
L53K1	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	C	N	C*5	C*7	C*4	C from first SN	C	C
L55		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C			C				C	N	C*5		C*4		C	C
L64		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C			C				C	N	C*5		C*4		C	C
MXS1		C							C		C	C			C	C	C																											C		C	N	N	N	N	—	N	N
MXS2ESLL		C									C	C					C													N													C			N		N			—		
MXS2ESLL1		C									C	C					C													N													C			N		N					
R41R		C							C		C	C	C		C*7	C*7	C	C	C						C				C	N	C						N				C		C		C	N	C*5	C*7	C*4	C from first SN	C	C
R41RL		C							C		C	C	C		C*7	C*7	C	C	C						C				C	N	C						N				C		C		C	N	C*5	N	C*4	C from first SN	C	C
S11	C	C					N		C		C	C	C	C	C	C	C								C	C	C	C	C	N	C	C	C	C	C		N	C	C			C	C	C	N	C*5	N	C*4	M00201～	C	C
S12		C							C		C	C	C		C	C	C	C	C						C				C	N	C						C				C		C		C	N	C*5	N	C*4	C from first SN	C	C
S121	C	C					N		C		C	C	C	C	C	C	C								C	C	C	C	C	N	C	C	C	C	C		C	C	C			C	C	C	N	N	N	C*4	C from first SN	C	C
S21		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C			C				C	N	C*5		C*4		C	C
S211		C							C		C	C	C		C	C	C	C	C						C				C	N	C						C				C		C		C	N	C*5	N	C*4	C from first SN	C	C

Hitachi Healthcare Americas

Name of Cleaning, Disinfection or Sterilization	3M Rapid Multi Enzyme Cleaner	CIDEZYME/ ENZOL	EMpower	Klenzyme	Metrizyme	Prolystica 2x conc. Enzymatic Presoak & cleaner	Transeptic	CaviWipes	CIDEX	Cidex ADS	CIDEX OPA	Cidex Plus	DETERGICIDE	Dynacide PA	EndoAct	EndoDis	HYAMINE	METRICIDE	METRICIDE 28	Metricide Plus 30	Metricide OPA Plus	NU-Cidex	Omnicide 14 N.S.	Omnicide 28	Oxivir Wipes	PI-SPRAY	PI-SPRAY II	PROTEX Disinfection Spary & Wipes	PROTEX ULTRA	Revital-Ox Resert XL HDL	Sani-Cloth Active Wipes	Sanicloth HB	Sani-Cloth Plus	SONO ULTRASOUND WIPES	Super Sani-Cloth Germicidal Wipe	TD100 & TD5	Trophon EPR	T-SPRAY	T-SPRAY II	Virkon	Virkon S	Wex-Cide 128	PERASAFE	Sporicidin	WAVICIDE-01	Autoclave (Steam)	EO sterilization	STERIS SYSTEM1	STERIS V-Pro	STERIS V-Pro Applicable SN	Sterrad 50/ 100S/ 200	Sterrad NX/ 100NX
S22	C	C					N		C		C	C	C	C	C	C	C								C	C	C	C	C	C	C	C	C		C	C	C	C			C	C	C	N	N	N	C*4	C from first SN	C	C
S31		C							C		C	C	C		C	C	C	C	C									C	C	C	C	C	C		C		C	C		C		C	N	C*5		C*4		C	C
S31KP	C	C					N		C		C	C	C	C	C*7	C*7	C								C	C	C	C	C	C	C	C	C	C	C		C	C	C			C	C	N	C*5	C*7	C*4	C from first SN	C	C
S3ESEL		C	C		C	C		N		C	C	C						C	C	C	C	C	C	C		C	C								C		N	C	C	C		N		C		C	N	N	N	—	N	—
S3ESL	C	C					N		C		C	C	C	C			C								N	C	C	C		C	C	N	C	C	C	C		N	C	C			C	N	C	C	N	N	—	N	—
S3ESL1		C					N				C	C					C								N							N						N			C	N		N	C*5	N	N	N	—	N	—
S3ESS	C	C					N		C		C	C	C	C			C								N	C	C	C		C	C	N	C	C	C	C		N	C	C			C	N	C	N	C*5	N	N	—	N	—
S42		C							C		C	C	C		C	C	C	C	C						C				C	N	C				C					C				C	N	C*5	N	C*4	C from first SN	C	C
SML44		C							C		C	C	C		C	C	C	C	C						C				C	N	C							C			C				C	N	C*5	N	C*4	C from first SN	C	N
VC34		C		C					C		C	C					C																				N			C				C	N	N	N	N	—	N	—
VC34A		C		C					C		C	C	N				N														N	N	N				N							N	N		N		N	N
VC35	C	C					N		C		C	C	C	C	C	C	C								N	C	C	C	C	C	C	N	C	C	C		N	C	C			C	C	C	N	N	N	N	—	N	—
VC41V	C	C					N		C		C	C	C	C	C	C	C								N	C	C	C	C	C	C	N	C	C	C		N	C	C			C	C	C	N	N	N	N	—	N	—
VL54		C		C					C		C	C					C																				N			C				C	N	N	N	N	—	N	—

NOTE

C: Compatible

C*1: Use with STERRAD® Booster (Part #15400) and STERRAD® Adapter II (Part #15402)

C*2: This product cannot be used with STERRAD NX.

C*3: This product cannot be used with STERRAD 50.

C*4: Compatible probes and accessories are tested for V-PRO 60, V-PRO maX, V-PRO 1 Plus and V-PRO 1.

C*5: Please refer to the instruction manual of the medical device for specific sterilization parameters.

C*6: Some probe housing and its accessories will be yellowed, the surface of some plastic materials become rough and the cable sheath becomes hard and tacky by Antigermix disinfector using UVC. Hitachi, Ltd. and Germitec confirmed that the function of the probe and its accessories do not suffer damages.

C*7: Optional waterproof case for probe connector is mandatory.

C*8: Trophon Curved Probe Positioner (CPP) is required.

N: No (do not use)

Blank: Not tested (do not use)